

Hema - screen 13

Numărător de celule



Numărător de celule semiautomat, cu dublă capilaritate, pentru determinarea simultană a 13 parametri hematologici:
WBC - RBC - Hgb - Hct - MCV - MCH - MCHC - Plt - MPV - Pct - RDW - PDW - LY%.
Histogramele WBC - RBC - Plt se afișează și se tipăresc.
Dilutor și imprimantă încorporate.
Software pentru analize veterinare.

CARACTERISTICI TEHNICE

DIMENSIUNE CAPILAR CITIRE: 72 micrometri pentru WBC, RBC și Plt
TIMP DE CITIRE: mai mic de 30"
VOLUMUL PROBEI: 20 μ L sânge integral
RAPORT DILUȚIE: RBC și Plt: 1:50.000, WBC 1:500
VOLUM ASPIRAȚIE: 250 μ L pentru RBC, WBC, Plt, 1500 μ L pentru Hgb
PRECIZIE:
WBC CV <2% (X=10.2x10³/ μ L)
RBC CV <1.5% (X=4.65x1000³/ μ L)
Hgb CV <1.2% (X=13.2 g/dL)
Plt CV <5% (X=250x10³/ μ L)
MCV CV <1% (X=87 μ m³)
REPETABILITATE: RBC și Hgb CV <1.5 – WBC și Hct CV=2 – Plt CV=3.5
ALIMENTARE: 220V a.c.; 50 Hz – 110V a.c.; 60 Hz
DIMENSIUNI: 36 (L) x 40 (H) x 42 (H) cm
GREUTATE: 15Kg.

```
RBC=4.21      WBC= 7.3
MCV= 94       HGB=12.5
HCT=39.7     PLT= 374
MCH=29.7     MPV= 9.1
MCHC=31.5    PCT=0.338
RDW=23.2%    PDW=32.5%
              LYMPH= 1.8%
Id= 2         thr= 27
```



All illustrations and specifications contained in this folder are based on the latest information available at the time of printing.
HOSPITEX DIAGNOSTICS reserves the right to make changes to materials, equipment and models, and to discontinue models at any time without notice.



**HOSPITEX
DIAGNOSTICS**



UNI EN ISO 9001

S.C. MENSANA IMPEX SRL
Str. Voltaire Nr.3, Ap.1
400130 CLUJ-NAPOCA

Str.Calea Mosilor, Nr.129
020854 BUCUREȘTI

Tel/Fax: 0264-597979,597391
E mail: hospitex@cluj.astral.ro
<http://hospitex.ro>; mensana.ro

Tel/Fax:021-3126536,031-1025604
E mail: hospitex@hospitex.ro